



Engage Your Patients in a Conversation About Dental Unit Waterlines

Q. I see you have information on 'Dental Infection Control Awareness Month (DICAM).' Is this a new thing?

A. It's an annual campaign held every September. The campaign highlights infection control in the dental office, which our office has always taken very seriously. DICAM reminds all of us how important it is to take these steps to keep you safe.

Q. I'm really concerned since I read about children getting infections from dental unit waterlines. Does this really happen?

A. There have been a few cases in which a small number of children became ill because of contaminated dental water. The dental community knows that if dental water isn't maintained and monitored properly, the health of patients is at risk. Let me tell you what we do in our office to keep our water safe (describe practice's protocol).

Q. I read a while ago that a dental unit waterline infection caused a patient to die. Is that true?

A. The dental unit waterline case you mention happened when an 80-year-old Italian patient acquired Legionnaire's disease after visiting her dentist for surgery. Unfortunately, the infection was traced back to an untreated dental unit waterline.

Q. What exactly is a dental unit waterline?

A. Dental unit waterlines are the long, narrow tubes that run to the dental chair (unit). These waterlines carry water from the public water supply to the dental equipment we use to treat and clean your teeth.

Q. How can the water be bad? Isn't it from the same place as the water we drink at home from the taps?

A. Yes, the water is from the public water supply. It's the long narrow tubes that are needed to supply water to our dental equipment that makes the difference. When water lingers in the tubing, or moves very slowly, germs can build-up inside the tubing and multiply. At home, the water we drink is from the same public water source, unless your home is supplied by a water well. Water supplied to our homes does not often linger and quickly streams through much wider pipes.

Q. What do you do to stop germs from building up in the dental unit waterline and causing infections?

A. We treat our dental unit waterlines on an ongoing basis with products specially designed for this purpose. We also test the water regularly as recommended by the CDC, to make sure that everything is working and that germs aren't able to build up in the tubing.

Q. What do you use to treat the water? Is it safe to use?

A. We follow the equipment manufacturer's instructions to treat the water safely and effectively. We use only treatment products that have been reviewed and tested, to be sure they are effective at preventing germs from building up in the waterlines but are also safe to be used around patients. These products are also reviewed by the Environmental Protection Agency (EPA) and/or Food and Drug Administration (FDA).

Q. Is it safe to use water from treated dental unit waterlines on me if I'm pregnant?

A. The treatments we use are safe for use for the general population. They are not contraindicated if you are pregnant.

For more information, please consult the following resources:

- Centers for Disease Control and Prevention. Guidelines for Infection Control in Dental Health-Care Settings — 2003. Available at: <http://www.cdc.gov/mmwr/PDF/rr/rr5217.pdf>.
- Centers for Disease Control and Prevention. Notes from the Field: *Mycobacterium abscessus* Infections Among Patients of a Pediatric Dentistry Practice — Georgia, 2015. Available at: <http://www.cdc.gov/mmwr/volumes/65/wr/mm6513a5.htm>.
- Centers for Disease Control and Prevention. Recommendations from the Guidelines for Infection Control in Dental Health-Care Settings — 2003. Available at: <http://www.cdc.gov/oralhealth/infectioncontrol/pdf/recommendations-excerpt.pdf>.
- Centers for Disease Control and Prevention. Summary of Infection Prevention Practices in Dental Settings. Basic Expectations for Safe Care. Available at: <http://www.cdc.gov/oralhealth/infectioncontrol/pdf/safe-care.pdf>.
- Organization for Safety, Asepsis and Prevention. Available at: www.osap.org.
- Ricci, ML, Fontana, S, Pinci, F, Fiumana E, Pedna MF, Farolfi P, Sabattini MA, Scaturro M. Pneumonia associated with a dental unit waterline. The Lancet. 2012;379(9816):684.



This resource was prepared by clinicians for the benefit of the Organization for Safety, Asepsis and Prevention (OSAP), a nonprofit, independent organization providing information and education on infection control and prevention and patient and provider safety to dental care settings worldwide. The information provided is for educational purposes only. It is not intended nor implied to be a substitute for the users' own exercise of their professional judgment. OSAP, and the clinicians who assisted it in preparing this material, expressly disclaims responsibility, and shall have no liability, for any damages, loss, injury, or liability whatsoever suffered as a result of your reliance on the information contained in this resource.